

Theme: Physics

Abstract No.: PTCOG-AO2025-ABS-0077

Abstract Title: Dosimetric Assessment of Bony Anatomy Position Variation in Prostate Stereotactic Body Proton Therapy (SBPT)

Author Names: Wing Ki Claudia Chan¹, Chin Chak HO¹, Hui GENG¹, Wai Wang LAM¹, Ka Keung TANG¹, Ting Chuan LI¹, Yeung Sum WONG¹, Shu Ting HUNG¹, Darren Ming Chun POON², Kin Yin CHEUNG¹, Siu Ki YU¹, Bin YANG¹

¹ Medical Physics Department, HKSH Eastern Medical Center, Hong Kong

² Comprehensive Oncology Centre, Hong Kong Sanatorium & Hospital, 2 Village Road, Happy Valley, Hong Kong

Background / Aims:

In prostate proton therapy (PT) utilizing a lateral opposing beam arrangement, a significant portion of proton beam passes through bony anatomy, such as femoral heads and pelvic bone. While target localization is based on fiducial markers matching, positional variation of prostate alters its spatial relationship with the surrounding bony anatomy. Current study simulated such positional changes between the bony anatomy and the prostate and assessed the dosimetric impact in prostate SBPT.

Subjects and Methods:

Six patients with localized prostate cancer that underwent 5-fraction SBPT were included. Treatment plans were optimized for 3mm patient positional uncertainty and $\pm 3\%$ range uncertainty (RU) using RayStation2024A TPS. To simulate the relative positional change between the bony anatomy and the prostate, a translational shift in each direction (right, left, inferior, superior, posterior or anterior) was applied to treatment plan iso-centre, contour of targets (CTV: prostate gland and GTV: intraprostatic lesion), and nearby soft tissues organs-at-risk (urethra, bladder, rectum and penile bulb), whereas the contour of body and bony structures (pelvic bone and femoral heads) remained unchanged. Six scenarios were assessed: 1-3) 5mm shift with and without $\pm 3\%$ RU, 4-6) 10mm shift with and without $\pm 3\%$ RU. Dosimetric evaluations were based upon clinical goals, with the occurrence of failed scenarios reported.

Result:

Table1 summarizes the scenario passing rate and mean clinical goal for failed scenarios of targets and OARs. The relative change in bony structure position slightly reduced target coverage as suggested by D95% and Dmin of CTV and GTV. However, the average D95% of CTV remained above 97% of prescription dose except 10mm shift with -3% RU. Apparently, bony structure position variations did not induce hot spots (D1cc or V105%) in target for most of the scenarios. Apart from D2% ≤ 28.5 Gy in penile bulb and Dmax ≤ 30 Gy in femoral heads, surrounding OARs (rectum and bladder) passed all clinical goals for all scenarios suggesting that dose spill over was minimal. For urethra inside the target, D0.5cc was occasionally larger than 40Gy. Figure1 shows that failed scenarios occur more frequently along LR direction. Plan iso-center shift along beam path has minimal effect on changing the dose distribution to match the target shift relative to bone and body.

Conclusion:

Dosimetric impact upon relative positional change between bony structure and prostate (up to 10mm) in prostate SBPT is minimal, demonstrating robustness in target coverage and OAR sparing across evaluated scenarios.




Figure 1. The occurrence of failed scenarios

	5mm Shift	5mm Shift, $\pm 3\%$ Range Uncertainty		5mm Shift, -3% Range Uncertainty		10mm Shift	10mm Shift, $\pm 3\%$ Range Uncertainty		10mm Shift, -3% Range Uncertainty			
Clinical Goals	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)	Passing Rate (Failed Scenarios)		
D95.0% ≥ 38 Gy (100%)	94.4%	84.4%	87.5% (98.2%)	84.4%	37.5% (98.2%)	66.7%	36.9% (97.1%)	81.1%	24.6% (99.5%)	50.0%	37.5% (98.2%)	
Dmax ≤ 28.5 Gy (100%)	83.3%	35.3% (83.0%)	85.0%	32.4% (88.0%)	100.0%	66.7%	31.9% (83.1%)	100.0%	25.4%	40.0% (99.8%)	63.9%	35.9% (97.2%)
D1.5cc ≤ 38.5 Gy (100%)	100.0%	100.0%	97.2% (86.0%)	100.0%	100.0%	100.0%	100.0%	100.0%	94.0%	48.0% (100.3%)	100.0%	100.0%
D2% ≤ 28.5 Gy (100%)	100.0%	87.2%	87.2%	7.5%	100.0%	100.0%	84.4%	100.0%	84.4%	2.5%	100.0%	100.0%
D95.0% ≥ 42.5 Gy (100%)	80.0%	42.2% (99.8%)	70.0%	41.5% (97.7%)	33.3%	41.5% (97.7%)	56.7%	40.2% (98.4%)	63.3%	39.3% (92.4%)	13.3%	40.3% (98.4%)
Dmax ≤ 40.375 Gy (95%)	100.0%	100.0%	100.0%	98.9% (93.3%)	100.0%	100.0%	98.3%	98.9% (98.8%)	73.3%	32.7% (98.7%)	88.9%	29.7% (95.3%)
V10.5% ≤ 0.4 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V48.8% ≤ 0.5 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V10.5% ≤ 0.5 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V22.2% ≤ 20.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V20.0% ≤ 10.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V42.2% ≤ 0.4 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V48.8% ≤ 2.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V27.0% ≤ 10.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V36.0% ≤ 40.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V20.0% ≤ 20.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
V18.0% ≤ 40.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Dose ≤ 60.0 Gy (81.1%)	45.5%	55.6%	48.7%	69.4%	40.7%	61.1%	41.5%	52.8%	48.9%	72.2%	41.0%	100.0%
D95.0% ≥ 38 Gy (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Dmax ≤ 38.5 Gy (100%)	100.0%	100.0%	100.0%	97.2%	33.6%	100.0%	100.0%	100.0%	100.0%	86.1%	33.0%	100.0%
V20.0% ≤ 10.0 cc (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Dmax ≤ 40.375 Gy (95%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
D2% ≤ 28.5 Gy (100%)	84.4%	28.8%	100.0%	83.3%	29.4%	88.9%	28.2%	100.0%	100.0%	83.3%	28.9%	100.0%
Dose ≤ 18.0 Gy (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Dmax ≤ 38.5 Gy (100%)	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%

Table 1. Summary of clinical goals in various scenarios